ICS 07. 060 N 92

# HY

Ocean Industry Standard of the People's

# Republic of China

HY/T 099-2007

# Test method of seawater nutrients

analyzer

# 海水营养盐测量仪检测方法

(English Translation)

Issue date: 2007-12-19

Implementation date: 2008-01-01

Issued by State Oceanic Administration of the People's Republic of China

# Foreword

SAC/TC 283 is in charge of this English translation. In case of any doubt about the contents of English translation, the Chinese original shall be considered authoritative.

Annex A of this standard is normative, while Annex B and Annex C are informative.

This standard was proposed by National Center of Ocean Standards and Metrology.

This standard was prepared by National Technical Committee 283 on Ocean of Standardization Administration of China (SAC/TC283).

# Test method of seawater nutrients analyzer

#### 1 Scope

This standard specifies the test items, test equipments, test environmental conditions, test methods of the seawater nutrients analyzer (hereinafter referred to as the analyzer) and writing requirements for test reports.

This standard is applicable to the test of seawater nutrients analyzer for long-term in-field use.

2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this standard. For dated references, subsequent amendments, or revisions, of any of these publications do not apply to this standard. However parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies.

HY 016.3 The basic method of environmental test of oceanographic instruments low temperature storage test

HY 016.4 The basic method of environmental test of oceanographic instruments high temperature test

HY 016.5 The basic method of environmental test of oceanographic instruments high temperature storage test

HY 016.8 The basic method of environmental test of oceanographic instruments alternate humidity test

HY 016.11 The basic method of environmental test of oceanographic instruments vibration test

HY 016.12 The basic method of environmental test of oceanographic instruments shock test

HY 016.13 The basic method of environmental test of oceanographic instruments bump test

HY 016.14 The basic method of environmental test of oceanographic instruments inclination and rolling-pitching test

HY 016.15 The basic method of environmental test of oceanographic instruments hydrostatic pressure test

1

#### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1 in-field seawater nutrients analyzer

Analyzer applicable to marine in-field investigation, and could inject sample, produce chemical reaction, analyze nutrients in seawater automatically Note: modify HY/T 093 - 2005, terms 3.2.

4 Technical requirements

4.1 Appearance

The appearance requirements for the analyzer are as follows:

- a) The shell of the analyzer, the paint coat and clad layer of the surface shall be uniform in color, smooth and firm, there shall be no obvious abrasion, rust corrosion, leakage, crack or bubble;
- b) The electrode lead connection of the analyzer shall be reliable, and the fasteners and connectors shall not be loose.
- 4.2 Metrological performance

The analyzer metrological performance requirements are given in table 1.

parameter	detection range / (µmol/L)	indication error / (µmol/L)
NO <sub>3</sub> -N	0.5~20	the reading value $\leqslant$ 5.0, $\pm$ 0.5 $\mu$ mol/L; the reading value $\geqslant$ 5.0, $\pm$ 10% $ imes$ reading value
NO <sub>2</sub> -N	0.5~7	the reading value $\leqslant$ 3.0, $\pm$ 0.3 $\mu mol/L$ the reading value $\geqslant$ 3.0, $\pm$ 10% $\times$ reading value
PO <sub>4</sub> -P	0.2~3	the reading value $\leqslant$ 1.3, $\pm$ 0.13 $\mu mol/L;$ the reading value $\geqslant$ 1.3, $\pm$ 10% $\times$ reading value
SiO <sub>3</sub> -Si	0.5~20	the reading value $\leqslant$ 5.0, $\pm$ 0.5 $\mu mol/L;$ the reading value $\geqslant$ 5.0, $\pm$ 10% $\times$ reading value
NH <sub>4</sub> -N	0.5~20	the reading value $\leqslant$ 5.0, $\pm$ 0.75 $\mu {\rm mol/L}$ ; the reading value $\geqslant$ 5.0, $\pm$ 15% $\times {\rm reading}$ value

Table 1 the analyzer metrological performance

5 Test items

The test items of the analyzer are as follows:

- a) Appearance inspection;
- b) Metrological performance tests include indication error, repeatability, temperature influence and detection limit;
- c) Environmental adaptability test.
- 6 Test equipments

The test equipments for metrological performance of analyzer are given in Table 2.

S/N	Equipments	Technical index					
1	nutrients standard	Certified Reference Material(CRM), or prepare nutrients					
1	substance	standard working solution as required in annex A					
2	analytical balance	minimum division value is 0.0001 g					
3	high-low temperature						
5	test chamber						
Note: Unless otherwise specified, the reagents used in this method are analytically pure (AR),							
and the water is deionized water or equivalent conductivity of pure water.							

Table 2 Test Equipments for Metrological Performance of analyzer

7 Test environmental conditions

The test environmental conditions requirements for the analyzer are as follows:

- a) Ambient temperature:  $(20\pm5)$  °C;
- b) Relative humidity :  $\leq$  80%;
- c) Supply voltage: AC  $(220\pm20)$  V,  $(50\pm2.5)$  Hz;
- d) There shall be no electromagnetic interference that affects the normal operation of the analyzer.
- 8 Test methods
- 8.1 Appearance inspection

Visual inspection and tactile examination shall be used to inspect the appearance of the analyzer according to 4.1.

8.2 Metrological performance test

8.2.1 NO<sub>3</sub>-N test

8.2.1.1 Indication error test

The NO<sub>3</sub>-N standard solution of primary CRM shall be used as measurement standard, and select 3 different concentrations of the standard solution to cover the measurement range. Or the NO<sub>3</sub>-N standard working solution prepared as required in Annex A.3 shall be used as the measurement standard.

At the ambient temperature of  $(20\pm5)$  °C, NO<sub>3</sub>-N standard solutions shall be measured according to the operation rules of the analyzer, in the process of venting, cleaning, self-inspection, sampling, and from low concentration to high concentration. The measurement of each standard solution shall be repeated for 3 times and the average value shall be calculated as measurement value of analyzer.

When NO<sub>2</sub>-N $\leq$ 3.0 µmol/L, NO<sub>3</sub>-N $\leq$ 5.0 µmol/L, NH<sub>4</sub>-N $\leq$ 5.0 µmol/L, PO<sub>4</sub>-P $\leq$ 1.3 µmol/L and SiO<sub>3</sub>-Si  $\leq$  5.0 µmol/L respectively, calculate indication error of the analyzer according to formula (1):

$$\Delta C_j = C_{jp} - C_{js} \qquad (1)$$

Where

 $riangle {\cal C}_{\it j}$  is the indication error on the j<sup>th</sup> concentration test point, (µmol/L);

 $C_{jp}$  is the average concentration measurement value on the j<sup>th</sup> concentration test point, (µmol/L);

 $C_{j_{s}}$  is the standard concentration value on the j<sup>th</sup> concentration test point, ( $\mu$ mol/L); When NO<sub>2</sub>-N $\geq$ 3.0  $\mu$ mol/L, NO<sub>3</sub>-N $\geq$ 5.0  $\mu$ mol/L, NH<sub>4</sub>-N $\geq$ 5.0  $\mu$ mol/L, PO<sub>4</sub>-P $\geq$ 1.3  $\mu$ mol/L and SiO<sub>3</sub>-Si  $\geq$  5.0  $\mu$ mol/L respectively, calculate indication error of the analyzer according to formula (2):

$$\delta = \left[ (C_{jp} - C_{js}) / C_{js} \right] \times 100\%$$
 (2)

#### Where

 $C_{jp}$  and  $C_{js}$  have the same meaning with formula (1);

 $\delta$  is the relative error of measurement; %.

8.2.1.2 Repeatability test

At the ambient temperature of  $(20\pm5)$  °C, any concentration of NO<sub>3</sub>-N standard solution shall be repeatedly measured no less than 6 times according to the operation rules of the analyzer.

Calculate the repeatability of the analyzer according to formula (3):

#### Where

 $\sigma$  is the experimental standard deviation of nutrients concentration measurement;  $C_i$  is the measurement concentration value on the i<sup>th</sup> concentration test point, ( $\mu$ mol/L);

 $C_{\rho}$  is the average value of the analyzer on the i<sup>th</sup> concentration test point, ( $\mu$ mol/L); *n* is measurement times.

#### 8.2.1.3 Temperature influence error

The analyzer and NO<sub>3</sub>-N standard solution of any concentration shall be placed in a high-low temperature test chamber, the chamber temperature shall be adjusted to  $5^{\circ}$ C and  $35^{\circ}$ C respectively, and analyzer and NO<sub>3</sub>-N standard solution shall be fully balanced to chamber temperature. According to the operation rules of the analyzer, the measurement of each standard solution shall be repeated separately for 3 times, and the average value shall be calculated as measurement value of analyzer. Separately calculate the error caused by temperature influence according to formula (4):

$$\Delta C_i = C_{i0} - C_{is} \quad \dots \quad (4)$$

Where

 $riangle \mathcal{C}_i$  is the error of indication on the i<sup>th</sup> temperature test point, (µmol/L);

 $C_{ip}$  is the average value on the i<sup>th</sup> temperature test point, ( $\mu$ mol/L);

 $C_{is}$  is the standard value of standard solution on the i<sup>th</sup> temperature test point, ( $\mu$ mol/L);

8.2.1.4 Detection limit test

At the ambient temperature of  $(20\pm5)$  °C, blank solution shall be measured standard repeatedly, according to the operation rules of the analyzer, and the measurement times shall be no less than 10. Calculate experimental standard deviation of analyzer measurement results according to formula (5):

Where

 $\sigma$  is the experimental standard deviation of blank solution, ( $\mu$ mol/L);

 $C_i$  is the measurement value of blank solution on the i<sup>th</sup> concentration test point, ( $\mu$ mol/L);

 $C_{P}$  is the average value of measurement results for blank solution, ( $\mu$ mol/L);

*n* is measurement times.

Calculate detection limit of the analyzer according to formula (6):

$$D.L = 2\sqrt{2} \cdot t \cdot \sigma \quad \dots \quad (6)$$

Where

*D.L* is the detection limit of analyzer,  $(\mu mol/L)$ ;

*t* is the bilateral *t*-distribution table (Refer to *t*-distribution table. When n=10, *a* (fiducial probability) =0.95, *t*=2.2281).

8.2.2 NO<sub>2</sub>-N test

8.2.2.1 Indication error test

The  $NO_2-N$  standard solution of primary CRM shall be used as measurement standard, and select 3 different concentrations of the standard solution to cover the measurement range. Or the  $NO_2-N$  standard working solution prepared as required in Annex A. 4 shall be used as the measurement standard. The test shall be conducted according to the method specified in 8.2.1.1.

8.2.2.2 Repeatability test

One of the above three NO<sub>2</sub>-N standard solutions shall be measured according to the method specified in 8.2.1.2.

8.2.2.3 Temperature influence error

One of the above three  $NO_2$ -N standard solutions shall be measured according to the method specified in 8.2.1.3.

8.2.2.4 Detection limit test

Blank solution shall be measured standard according to the method specified in 8.2.1.4.

8.2.3 PO<sub>4</sub>-P test

8.2.3.1 Indication error test

The PO<sub>4</sub>-P standard solution of primary CRM shall be used as measurement standard, and select 3 different concentrations of the standard solution to cover the measurement range. Or the PO<sub>4</sub>-P standard working solution prepared as required in Annex A.5 shall be used as the measurement standard.

The test shall be conducted according to the method specified in 8.2.1.1.

8.2.3.2 Repeatability test

One of the above three PO<sub>4</sub>-P standard solutions shall be measured according to the method specified in 8.2.1.2.

8.2.3.3 Temperature influence error

One of the above three  $PO_4$ -P standard solutions shall be measured according to the method specified in 8.2.1.3.

8.2.3.4 Detection limit test

Blank solution shall be measured standard according to the method specified in 8.2.1.4.

8.2.4 SiO<sub>3</sub>-Si test

8.2.4.1 Indication error test

The SiO<sub>3</sub>-Si standard solution of primary CRM shall be used as measurement standard, and select 3 different concentrations of the standard solution to cover the measurement range. Or the SiO<sub>3</sub>-Si standard working solution prepared as required in Annex A.6 shall be used as the measurement standard.

The test shall be conducted according to the method specified in 8.2.1.1.

8.2.4.2 Repeatability test

One of the above three  $SiO_3$ -Si standard solutions shall be measured according to the method specified in 8.2.1.2.

8.2.4.3 Temperature influence error

One of the above three SiO<sub>3</sub>-Si standard solutions shall be measured according to the method specified in 8.2.1.3.

8.2.4.4 Detection limit test

Blank solution shall be measured standard according to the method specified in 8.2.1.4.

8.2.5 NH<sub>4</sub>-N test

8.2.5.1 Indication error test

The NH<sub>4</sub>-N standard solution of primary CRM shall be used as measurement standard, and select 3 different concentrations of the standard solution to cover the measurement range.

Or the  $NH_4$ -N standard working solution prepared as required in Annex A.7 shall be used as the measurement standard.

The test shall be conducted according to the method specified in 8.2.1.1.

8.2.5.2 Repeatability test

One of the above three  $NH_4$ -N standard solutions shall be measured according to the method specified in 8.2.1.2.

8.2.5.3 Temperature influence error

One of the above three  $NH_4-N$  standard solutions shall be measured according to the method specified in 8.2.1.3.

8.2.5.4 Detection limit test

Blank solution shall be measured standard according to the method specified in 8.2.1.4.

8.3 Environmental adaptability test

8.3.1 Environmental test

The specific parameters of environmental test items shall be determined according to the environmental conditions of the analyzer. The environmental test methods are given in Table 3.

7

S/N	Environmental test item	Environmental test method
1	Low temperature storage test	According to HY 016.3
2	Low temperature test	According to HY 016.2
3	High temperature storage test	According to HY 016.5
4	High temperature test	According to HY 016.4
5	Shock test	According to HY 016.12
6	Inclination and rolling-pitching test	According to HY 016.14
7	Bump test	According to HY 016.13
8	Vibration test	According to HY 016.11
9	Alternate humidity test	According to HY 016.8
10	Hydrostatic pressure test	According to 016.15

#### Table 3 environmental test method of analyzer

8.3.2 Verification test of metrological performances

After the environmental test, the metrological performances shall be tested again according to 8.2.

9 Test Report

The test report shall be accurately and objectively, and each test result shall include the following information:

a) The title -"Test Report";

b) Name and address of test organizations, place where test carried out;

c) Unique identification of the test report and its each page, for ensuring that the page is the part of the test report; and clear identification indicating the end of the test report;

d) Customer's name and address;

e) Tested analyzer name, model/type, serial number, and manufacturer;

f) Technical documentations to be based on;

g) Name, position, signature or equivalent identification of authorizer of the test report;

h) Signature or equivalent identification of analyst and verifier;

i) Description and state of the tested instrument;

j) List of main measuring instruments used in the test, including model/type, uncertainty /accuracy /maximum permissible error, certificate Number.

k) Test time, location and environmental conditions;

I) Test results.

The test record format see Annex B.

The test report format see Annex C.

#### Annex A

#### (Normative Annex)

#### Preparation of Nutrients Standard Solution

#### A.1 Reagent materials

Unless otherwise specified, the chemical reagents used in this standard shall be analytical reagent (Grade AR), and the water shall be deionized or equivalent conductivity of pure water.

A.2 Preparations of artificial seawater

Weigh 31 g NaCl (Grade GR), 10 g MgSO<sub>4</sub>•7H<sub>2</sub>O (Grade GR), 0.05 g NaHCO<sub>3</sub>•H<sub>2</sub>O (Grade GR) and dissolve in water. Diluted to 1 L with the deionized water. The salinity is 35. A.3 Preparation of NO<sub>3</sub>-N standard solutions

A.3.1 NO<sub>3</sub>-N standard stock solutions

Weigh 1.011 g KNO<sub>3</sub> (pre-dried at 110° C for 1 h, then cooled to room temperature in the drying vessel) and dissolve in a little of deionized water. Dilute to 1000 mL volumetric flask, and add 1.0 mL CHCl<sub>3</sub>. The concentration is 10.00 mmol/L NO<sub>3</sub>-N, and its valid period is 6 months.

A.3.2  $NO_3$ -N standard working solution

Take 10.00 mL standard stock solutions and dilute to 100 mL volumetric flask by deionized water and mix evenly. The solution is 1.0mmol/L NO<sub>3</sub>-N, which is prepared before use.

A.3.3 NO<sub>3</sub>-N standard solution

Take 1.00 mL, 5.00 mL, 20.0 mL, 50.0 mL and 90.0 mL NO<sub>3</sub>-N standard working solution separately, dilute to five 1 000 mL volumetric flasks with artificial seawater and mix evenly. The concentration is respectively 1.0  $\mu$ mol/L, 5.0  $\mu$ mol/L, 20.0  $\mu$ mol/L, 50.0  $\mu$ mol/L.

A.4 Preparation of  $NO_2$ -N standard solution

A.4.1  $NO_2$ -N standard stock solution

Weigh 0.3450 g NaNO<sub>2</sub> (pre-dried at 110°C for 1 h, then cooled to room temperature in the drying vessel) and dissolve with a little of deionized water, then, dilute to 500 mL volumetric flask, and add 1.0 mL CHCl<sub>3</sub>. The concentration is 10.00 mmol/L NO<sub>2</sub>-N, and its valid period is 2 months.

A.4.2 NO<sub>2</sub>-N standard working solution

Take 10.00 mL standard stock solutions and dilute to 100 mL volumetric flask by deionized water and mix evenly. The solution is 1.0mmol/L NO<sub>3</sub>-N, which is prepared

before use, and its valid period is 4h.

A.4.3  $NO_2$ -N standard solution

Take 1.00 mL, 5.00 mL, 20.0 mL, 50.0 mL and 90.0 mL NO<sub>2</sub>-N standard working solution separately, dilute to five 1 000 mL volumetric flasks with artificial seawater and mix evenly. The concentration is respectively 1.0  $\mu$ mol/L, 5.0  $\mu$ mol/L, 20.0  $\mu$ mol/L, 50.0  $\mu$ mol/L.

A.5 Preparation of PO<sub>4</sub>-P standard solution

A.5.1 PO<sub>4</sub>-P standard stock solution

Weigh 1.088 g KH<sub>2</sub>PO<sub>4</sub> (pre-dried at 110° C for 1 h, then cooled to room temperature in the drying vessel) and dissolve with a little of deionized water, then, dilute to 1000 mL volumetric flask, and add 1.0 mL CHCl<sub>3</sub>. The concentration is 8.00 mmol/L PO<sub>4</sub>-P, and its valid period is 6 months.

A.5.2 PO<sub>4</sub>-P standard working solution

Take 10.00 mL standard stock solutions and dilute to 100 mL volumetric flask by deionized water and mix evenly. The solution is 0.8mmol/L PO<sub>4</sub>-P, which is prepared before use, and its valid period is 24h.

A.5.3 PO<sub>4</sub>-P standard solution

Take 1.00 mL, 3.00 mL, 5.00 mL PO<sub>4</sub>-P standard working solution separately, dilute to five 1 000 mL volumetric flasks with artificial seawater and mix evenly. The concentration is respectively 0.8  $\mu$ mol/L, 2.4  $\mu$ mol/L and 4.0  $\mu$ mol/L.

A.6 Preparation of SiO<sub>3</sub>-Si standard solution

A. 6. 1 SiO<sub>3</sub>-Si standard stock solution

Weigh 4.702 g Na<sub>2</sub>SiF<sub>6</sub> (pre-dried at 110°C for 1 h, then cooled to room temperature in the drying vessel) and dissolve with a little of deionized water, then, dilute to 1000 mL volumetric flask, and add 1.0 mL CHCl<sub>3</sub>. The concentration is 25.0 mmol/L SiO<sub>3</sub>-Si, and its valid period is 12 months.

A. 6. 2 SiO<sub>3</sub>-Si standard intermediate solution

Take 8.00 mL standard stock solutions and dilute to 200 mL volumetric flask by deionized water, add 3 drops of CHCl<sub>3</sub> and mix evenly. The solution is 1.00mmol/L SiO<sub>3</sub>-Si, and its valid period is 24h.

A.6.3 SiO<sub>3</sub>-Si standard working solution

Take 10.00 mL standard intermediate solution and dilute to 100 mL volumetric flask by deionized water and mix evenly. The solution is 0.100mmol/L SiO<sub>3</sub>-Si, which is prepared before use.

A.6.4 SiO<sub>3</sub>-Si standard solution

Take 5.00 mL, 20.0 mL SiO<sub>3</sub>-Si standard working solution and 20.0mL, 50.0mL, 80.0mL SiO<sub>3</sub>-Si standard intermediate solution separately, dilute to five 1 000 mL volumetric flasks with artificial seawater and mix evenly. The concentration is respectively 0.50  $\mu$ mol/L, 2.00  $\mu$ mol/L, 20.0  $\mu$ mol/L, 50.0  $\mu$ mol/L, and 80.0  $\mu$ mol/L.

A.7 Preparation of  $NH_4$ -N standard solution

A.7.1  $NH_4\text{--}N$  standard stock solution

Weigh 0.5349 g NH<sub>4</sub>Cl (pre-dried at 110°C for 1 h, then cooled to room temperature in the drying vessel) and dissolve with a little of deionized water, then, dilute to 1000 mL volumetric flask, and add 1.0 mL CHCl<sub>3</sub>. The concentration is 10.0 mmol/L NH<sub>4</sub>-N, and its valid period is 6 months.

A.7.2 NH<sub>4</sub>-N standard intermediate solution

Take 20.00 mL standard stock solutions and dilute to 200 mL volumetric flask by deionized water, and mix evenly. The solution is 1.00 mmol/L NH<sub>4</sub>-N, and its valid period is at current day.

A.7.3 NH<sub>4</sub>-N standard working solution

Take 10.00 mL standard stock solutions and dilute to 100 mL volumetric flask by deionized water, and mix evenly. The solution is 0.100 mmol/L  $NH_4$ -N, which is prepared before use.

A.7.4 NH<sub>4</sub>-N standard solution

Take 5.00 mL, 20.0 mL NH<sub>4</sub>-N standard working solution and 20.0mL, 50.0mL, 80.0mL NH<sub>4</sub>-N standard intermediate solution separately, dilute to five 1 000 mL volumetric flasks with artificial seawater and mix evenly. The concentration is respectively 0.50 µmol/L, 2.00 µmol/L, 20.0 µmol/L, 50.0 µmol/L and 80.0 µmol/L.

# Annex B

# (Informative Annex)

## Formats of test record

### The formats of test record are given in Table B.1 and Table B.2.

### Table B.1 Formats of Analyzer Metrological Performance Test

Custor	ner										
Model/						Seria	l No.				
Main measuring instruments used in the test						·					
Test loc	ation						Test d	date			
Environmental	conditions	Tempera	ture:		°C	Hum	idity	: %F	ЯH		
Appeara	ance										
Standard cond	centration		Insti	rument	t indicat	ion	µmol∕L	_		Indicatio	n error
µmol∕	/L	1			2		3	Avera	age	µmol,	/L
		Instrument indication Wmol/I					Experim	ental			
Standard cond µmol/	Standard concentration µmol/L		2	3	4	5	6	Avera	age	standard deviatio	
Solution	Standard	Ins	Instrument indication µmol/L								
concentra temperature °C µmol/L		1		2			3 Average		ige	Indication µmol,	n error /L
Standard concentration µmol/L		Ins	trument	indi	cation $\mu$	mol/L	-	I		Experimental standard deviation µmol/L	Detection limit µmol/L
	1	2	3	3	4		5	ė	)	_	
			<u> </u>		10					-	
	/	8	, · · ·	1	10		11	1	Z	-	

Tested by

Checked by

Instrument name	Test item
Test date	Reference of the test
Room temperature	Relative humidity
Test conditions	
Test record	
Tested by	Checked by

Table	B. 2	Formats	of	instrument	environmental	test	
		1 01 1110 00	•.				

## Annex C

## (Informative Annex)

## Formats of test report

The front-page format of test report is given in Figure C.1, and the inside page format is given in Figure C.2 and C.3.

# XXX (Organization name)

# Test Report

I	Report No
Customer	
Customer Address	
Instrument name	
Model/Specification	
Serial Number	
Manufacturer	
lssue date	

(Stamp for Test)

Approved by <u>(Signature)</u>

Name (typed in Roman)

Position \_\_\_\_\_

Address:

Postal code:

Fax:

Tel.:

		Μ	lain mea	suring	instrum	ents use	ed in th	ne tes	t			
N	Model/Type				Uncertainty / Accuracy / Maximum Permissible Error			Certificate No.				
		Test	time,	Locatio	n, and l	Environ	mental c	condit	ions			
Location:							D	ate:				
Temperatur	re:		Relative humidity:									
			Test re	sults f	or metro	ologica	l perfor	mance			1	
Standar	d value			Instr	ument i	ndicati	on µmol/	/L			Indication	
μmo	I/L		1		2	3			average		error µmol∕L	
			Experimental									
Standar	d value	1	2	2 4		E		4	4		standard	
PHIO	1/L			5 4		5		0	o avera			
											Philo 17 E	
Solution				Instr	ument i	ndicati	on µmol/	/L			Indication	
temper-a ture ℃	Standar dµmol∕L		1		2		3		average	9	error µmol∕L	
		Exper							perimen	imental		
Blank s	olution		Instru	ument in	dicatio	on µmol∕L			standard		Detection	
μmo1/L								d	deviation		limit µmol∕L	
		1	2	2	4	F	4		µmoi/L			
		I	2	3	4	5	0					
		7	Q	0	10	11	12	-				
	/	0	,	10		12	_					
			1	1	1	1	1				1	
<b>T</b>						0						

# Figure C.1 Homepage format of test report

Tested by

Checked by

# Figure C.2 Inside page format of test results for metrological performance of

Main measuring instruments used in the test									
Name	Model/Type	Uncertainty / Accuracy /Maximum Permissible Error	Certificate No.						
	Test time, location, and e	environmental conditions							
Location:		Date:							
Temperature:		Relative Humidity:							
	Environmental adaptał	pility test results							

test report

Tested by

Checked by

Figure E.3 Inside page format of test results for environmental adaptability of test

report

References:

[1] GB 17378.4 The Specification for Marine Monitoring Part 4: Seawater Analysis.

[2] GB/T 12763.4-1991 The specification for oceanographic survey observations of chemical parameters in seawater.

[3] HY/T 008-1992 Marine analyzers terminology.

[4] HY/T 093-2005 *Seawater nutrients automatic analyzer*.